AVAILABLE CYCLES *(Complete list in the user manual)*

Cycle 1
Simple cleaning

- Leak test
- Pre-Cleaning
- Cleaning
- Rinsing
- Disinfect. 5 min.
- Rinsing
- Purging***

TIME*: 17 min.

Cycle 2
Double cleaning

- Leak test
- Pre-Cleaning
- Cleaning
- Rinsing
- Cleaning
- Rinsing
- Disinfect. 5 min.
- Rinsing
- Purging***
- Alcohol inject.**
- Purging***

TIME* : 19 min.

Cycle 3
With double disinfectant dose and 15' contact time

- Leak test
- Pre-Cleaning
- Cleaning
- Rinsing
- Cleaning
- Rinsing
- Disinfection 2xdoses 15min.
- Rinsing
- Purging***

TIME:* 35 min.

Cycle 4
Disinfection without cleaning

- Leak test
- Pre-Cleaning.
- Disinfect. 5 min.
- Rinsing
- Purging***

TIME* : 18 min.

Cycle 5
Additional purging

- Purging***
- Alcohol inject.**
- Purging***

TIME* : 5 min.

Cycle 6
"Selfdisinfection" After each daily program and before if the Soluscope has not been used for 72H

- Machine + 0.2µm Filter Chlorine Disinfection
- Rinsing
- Machine + 0.2µm Filter Glutaraldehyde Disinfection
- Draining

STAND BY
Press "ENTER" to reuse the machine

- Rinsing
- Purging***

TIME* : 45min

Cycle 8
"Descaling"

- Machine + 0.2µ Filter Descaling
- Rinsing
- Machine + 0.2µ Filter Glutaraldehyde Disinfection
- Draining
- Purging***

TIME* : 40'

THE INSTRUCTIONS ON THIS PAGE DO NOT EXEMPT THE USER FROM READING THE USER MANUAL

* Cycles times may vary according to the type of endoscope and the filters clogging (+/- 2min.)
** If alcohol option is present and activated.
*** The Purging is enough for an immediate reuse of the endoscope. For a long time storage, have the endoscope manually dried with medical air.

LAUNCHING A CYCLE

Right after the endoscopy, the endoscope must always be pretreated before being loaded into the Soluscope. A minimum pretreatment protocole is presented in the user manual

Select the correct connectics and connect the endoscope according to its No.
(See doc « Connectics choice »)

→ **Cycle 6**: connect And pour 12 tablets of bleach at 1.5g of active chlorin into the bowl.

→ **Cycle 8**: connect And pour 1x dose of Soluscope descaling product into the bowl.

Close the lid without forcing

From ground menu (Press "ESC" to display it) .



Press "ENTER" and follow carefully the instructions on display panel.

- **To move** the cursor, **press on direction kees (→)**
- **To cancel** the data input, press **"ESC"**
  To interrupt the cycle in progress, press **"ESC" , then +/-** for 3 seconds